

DURRIDGE COMPANY Inc


524 Boston Road, Billerica, Massachusetts 01821 Phone: 978.667.9556 • Fax: 978.667.9557 Web: www.durridge.com • Email: info@durridge.com

ROCK SAMPLE RADON CALIBRATION

Rock Sample SN:	Average of Four Readings:	:pCi/L
		Bq/M³
with dry, ambient air for not less than 10 standard rock-sample tubing set and sm setting reduced to 06. The RAD7 was so valves on the rock sample were opened aball valves were closed and the date of clast four were averaged. The data is present the RAD7 Cal. Date (Y/M/D): RAD7 Serial Number: RAD7 RH:	an one month. A recently calibrated, stand on minutes. The RAD7 was connected to all drying tube. The RAD7 protocol was witched OFF then ON and allowed to proud a test started. After completion of the consure noted on the rock sample. Of the cented below. Ambient Radon CC	the rock sample with a is set to 1-day and the 'Recycle' rint out a header. The ball he run (three hours later), the e six half-hour readings, the
	n will give the same result to within 7% actly the same setup and procedure, as c	
	Signature: Date (Y/M/D):	
	Date (1/11/1/D)	


Notes on Rock Sample Calibration

1. Natural Rock Sample Components Used:



3. Analysis Using CAPTURE Software

(Last four records examined and averaged):



4. Non-Standard Setups

If Vsetup is the total volume of the standard setup and Rav is the measured average concentration with the standard setup, the total radon in the system is Rav * Vsetup. For a non-standard setup with a different total volume, the average radon concentration of the four readings, RavNew, will be:

RavNew = Rav * (Vsetup / VsetupNew)

The standard setup component volumes are: Rock Sample 460ml, tubing set 4ml, small drying tube 24ml and RAD7 800ml. Thus the total standard setup volume, Vsetup, is 1,260ml.